

BM / HITACHI 912

04-01 CHEMISTRY PARAMETERS

CK-MB FS

TEST [CKMB] [#]	TEST NAME [CKMB]	UNIT [U/1]					
DATA MODE [ON BOARD]	REPORT NAME [CK-MB]						
CONTROL INT [0]	INSTR. FACT. (Y=aX + b) a [1.0]	b [0.0]					
EXPECTED VALUE	CLASS 1	EXPECTED VALUE CLASS 2					
AGE	M	F					
[] [] []	- [] []	- [] []					
[] [] []	- [] []	- [] []					
[0]	- [24]	[0] - [24]					
TECHNICAL LIMIT	CLASS I	CLASS 2					
[0]	- [1000]	[] - []					
STD	CON	POS	S.VOL	PRE. DIL.	VOL	CODE LOT	QUALITATIVE
							[NO]
(1)	[0.0][#]	[10]	[0]	[0]	[#]		(1) [] []
(2)	[*] [#]	[10]	[0]	[0]	[#]		(2) [] []
(3)	[] [] []	[] [] []	[] [] []	[] [] []	[] [] []		(3) [] []
(4)	[] [] []	[] [] []	[] [] []	[] [] []	[] [] []		(4) [] []
(5)	[] [] []	[] [] []	[] [] []	[] [] []	[] [] []		(5) [] []
(6)	[] [] []	[] [] []	[] [] []	[] [] []	[] [] []		(6) [] []

Order information

Cat. No.	Kit size
10 165 021	R1 5 x 20 ml + R2 1 x 25 ml
10 165 022	R1 5 x 80 ml + R2 1 x 100 ml
10 165 023	R1 1 x 800 ml + R2 1 x 200 ml
10 165 700	R1 16x 20 ml + R2 4 x 20 ml

Notes

- Please refer to the package insert for CK-MB FS for the detailed information about the test on the following:

Clinical Relevance
 Method and Principle
 Composition and Stability of the Reagents
 Specimens
 Calibrators and Controls
 Performance Characteristics concerning;
 Measuring Range
 Specificity/Interferences
 Sensitivity/Limit of Detection
 Precision (Reproducibility,
 Repeatability)
 Method Comparison
 Reference Ranges
 Literature

- The stability of the reagent on board of the analyser is at least one month provided that contamination and evaporation are avoided.
- Manufactured by
 DiaSys Diagnostic Systems GmbH & Co.KG
 Alte Strasse 9, 65558 Holzheim, Germany.

TEST [CKMB]
ASSAY CODE [Rate A] [10] [] WAVELENGTH (SUB / MAIN) [415] / [340]
ASSAY POINTS [25] - [31]- [0] - [0] DILUTION [301][99] < CLASS 1 > < CLASS 2 >
S.VOL (NORMAL) [10] [0] [0] [] [] []
S. VOL (DECREASE) [5] [0] [0] [] [] []
S.VOL (INCREASE) [20] [0] [0] [] [] []
ABS. LIMIT [9000] [] [2:INCREASE]
PROZONE LIMIT [0] [] [LOWER]
REAGENT R1 [200] [0] [#] [#]
R2 [0] [0] [#] [#]
R3 [50] [0] [#] [#]
R4 [0] [0] [#] [#]
CALIB. TYPE [1:LINEAR] [2] [2] [0] []
AUTOCALIB.
TIME OUT BLANK [0] SD LIMIT [0.1]
SPAN [0] DUPLICATE LIMIT [200]
2 POINT [0] SENSITIVITY LIMIT [0]
FULL [0] SI ABS. LIMIT [-32000][32000]
CHANGE LOT [NO] COMPENSATED LIMIT []
BOTTLE [NO]

Data entry by the user
 * Enter calibration or standard value
 ** The given factor must be checked by a calibration serum.
 ## Enter the next code