

HITACHI 717

CHLORIDE FS

Chemistry settings

Temperature: 37 °C

PROGRAM 2: CHEMISTRY PARAMETERS

TEST	CL
ASSAY CODE	1 - 50 - 0
SAMPLE VOLUME (µl)	3
R 1 VOLUME (µl)	250 - 100 - NO
R 2 VOLUME (µl)	0 - 20 - NO
WAVELENGTH (nm)	600 - 480
CALIBR. METHOD	1 - 0 - 0
STD.1 CONC.-POS.	0.0 - 1
STD.2 CONC.-POS.	* - *
STD.3 CONC.-POS.	0 - 0
STD.4 CONC.-POS.	0 - 0
STD.5 CONC.-POS.	0 - 0
STD.6 CONC.-POS.	0 - 0
SD LIMIT	0.1
DUPLICATE LIMIT	100
SENSITIVITY LIMIT	0
ABS.LIMIT(INC/DEC)	0 - 0
PROZONE LIMIT	0
EXPECTED VALUE	95 - 105
PANIC VALUE	1 - 130
INSTRUMENT FACTOR	1.0

#) Data entry by the user

*) Enter calibration or standard value and position

Order information

Cat. No.	Kit size
10 120 021	R 5 x 25 ml + 1 x 3 ml Std
1 1200 99 10 026	R 6 x 100 ml
10 120 023	R 1 x 1000 ml
10 120 030	6 x 3 ml Standard
10 120 717	R 6 x 100 ml

Notes

1. Please refer to the package insert for Chloride FS for detailed information about the test on the following:

Clinical Relevance
Method and Principle
Composition and Stability of the Reagents
Specimens
Calibrators and Controls
Performance Characteristics regarding
- Measuring Range
- Specificity/Interferences
- Sensitivity/Limit of Detection
- Precision (Reproducibility, Repeatability)
- Method Comparison

Reference Ranges
Literature

2. The stability of the reagent on board the analyser is at least one month provided that contamination and evaporation are avoided.
3. Manufactured by
DiaSys Diagnostic Systems GmbH & Co.KG
Alte Strasse 9, 65558 Holzheim, Germany