

Quality Control

## Certificate of Analysis

QC release date: 2018-01-16

**Product:** Albumin FS  
**Product Code:** 10220

Cat. No.: 1 0220 99 10 923

---

**Delivery of kits:** No of kits: 6  
Kit Lot/Los: 60124262  
Kit Expiry: 2020-01

---

**Delivery of bulk:** Volume:  
Processing Lot:

---

### Results for:

Reagent Lot No.: 24777  
Expiry: 2020-01

### Examination:

Visual check: Reagent : clear, pale green  
Smell Reagent : no characteristic odour  
**pH:** Reagent ( 25 °C ± 1 °C): 4.23

**Absorbance:** Reagent : 0.248  
546 nm/ 15-25°C

<b>Control sera:</b>	TruLab N	TruLab P
	Lot: 22255	Lot: 22256
test results	3.99 g/dL	5.21 g/dL
Target values	3.86 2.97 - 4.75	4.91 3.78 - 6.04

---

### Additional remarks:

This lot meets our specifications.  
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka