

Quality Control

Certificate of Analysis

QC release date: 2018-04-04

Product: Albumin FS
Product Code: 10220

Cat. No.: 1 0220 99 10 923

Delivery of kits: No of kits: 10
Kit Lot/Los: 60124883
Kit Expiry: 2020-03

Delivery of bulk: Volume:
Processing Lot:

Results for:

Reagent Lot No.: 25026
Expiry: 2020-03

Examination:

Visual check: Reagent : clear, pale green
Smell Reagent : no characteristic odour
pH: Reagent (25 °C ± 1 °C): 4.24

Absorbance: Reagent : 0.257
546 nm/ 15-25°C

Control sera:	TruLab N	TruLab P
	Lot: 22255	Lot: 22256
test results	3.92 g/dL	5.03 g/dL
Target values	3.86 2.97 - 4.75	4.91 3.78 - 6.04

Additional remarks:

This lot meets our specifications.
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka