

Quality Control

Certificate of Analysis

QC release date: 2019-04-16

Product: Creatinine PAP FS

Product Code: 11759

Cat. No.: 1 1759 99 10 026

Delivery of kits: No of kits: 6
Kit Lot/Los: 60132448
Kit Expiry: 2020-04

Delivery of bulk: Volume:
Processing Lot:

Results for:

Reagent 1 Lot No.: 27077
Expiry: 2020-04

Reagent 2 Lot No.: 27077
Expiry: 2020-04

Examination:

Visual check Reagent 1: clear, yellow
Reagent 2: clear, pale brown

Smell Reagent 1: like detergent
Reagent 2: no characteristic odour

pH: Reagent 1 (25 °C ± 1 °C): 8.09
Reagent 2 (25 °C ± 1 °C): 8.03

Absorbance: Reagent 1: 0.004
546 nm / 37 °C R1 + R2: 0.010

Control sera: TruLab N TruLab P
Lot: 24382 Lot: 24383

Test results 1.50 mg/dL 4.90 mg/dL

Target values 1.46 1.14 - 1.78 4.88 3.81 - 5.95

Additional remarks: Mf-Date: 13.02.2019

This lot meets our specifications.
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka