

Quality Control

Certificate of Analysis

QC release date: 2021-06-24

Product: Creatinine PAP FS

Product Code: 11759

Cat. No.: 1 1759 99 10 920

Delivery of kits: No of kits: 2
Kit Lot/Los: 60148589
Kit Expiry: 2022-06-30

Delivery of bulk: Volume:
Processing Lot:

Results for:

Reagent 1 Lot No.: 30924
Expiry: 2022-06-30

Reagent 2 Lot No.: 30924
Expiry: 2022-06-30

Examination:

Appearance: Reagent 1: clear, yellow
Reagent 2: clear, pale brown

Odour: Reagent 1: like detergent
Reagent 2: no characteristic odour

pH: Reagent 1 (25 °C ± 1 °C): 8.05
Reagent 2 (25 °C ± 1 °C): 8.01

Absorbance: Reagent 1: 0.005
546 nm / 37 °C R1 + R2: 0.010

Control sera: TruLab N TruLab P
Lot: 27547 Lot: 27548

Test results 1.39 mg/dL 5.19 mg/dL

Target values 1.39 1.08 - 1.70 5.15 4.02 - 6.28

Additional remarks: Mf-Date: 27.04.2021

This lot meets our specifications.
This certificate of analysis is a computer printout and has therefore not been signed
by hand.

Sign.: Dr. G. Gorka