

Quality Control

Certificate of Analysis

QC release date: 2018-06-13

Product: Uric Acid FS (TOOS)

Product Code: 13001

Cat. No.: 1 3001 99 10 920

Delivery of kits: No of kits: 8
Kit Lot/Los: 60126250
Kit Expiry: 2020-06

Delivery of bulk: Volume:
Processing Lot:

Results for:

Reagent 1 Lot No.: 25141
Expiry: 2020-06

Reagent 2 Lot No.: 25141
Expiry: 2020-06

Examination:

Visual check Reagent 1: clear, colourless
Reagent 2: clear, yellow to pale brown

Smell Reagent 1: no characteristic odour
Reagent 2: no characteristic odour

pH: Reagent 1 (25 °C ± 1 °C): 6.98
Reagent 2 (25 °C ± 1 °C): 6.98

Absorbance: Reagent 1: 0.002
546 nm / 15-25 °C R1 + R2: 0.020

Control sera: TruLab N TruLab P
Lot: 22255 Lot: 22256

Test results 6.77 mg/dL 7.16 mg/dL

Target values 6.63 5.70 - 7.56 7.12 6.12 - 8.12

Additional remarks: Mf-Date: 30.05.2018

This lot meets our specifications.
This certificate of analysis is a computer printout and has therefore not been signed
by hand.

Sign.: Dr. G. Gorka