

Quality Control

## Certificate of Analysis

QC release date: 2018-11-14

**Product:** Alkaline phosphatase FS ( IFCC 37 °C )

**Product Code:** 10441

Cat. No.: 1 0441 99 10 021

---

**Delivery of kits:** No of kits: 50  
Kit Lot/Los: 60128946  
Kit Expiry: 2020-02

---

**Delivery of bulk:** Volume:  
Processing Lot:

---

### Results for:

Reagent 1 Lot No.: 25888  
Expiry: 2020-02

Reagent 2 Lot No.: 25888  
Expiry: 2020-02

### Examination:

Visual check Reagent 1: clear, colourless liquid  
Reagent 2: clear, yellowish liquid

Smell Reagent 1: slight smell ( amines )  
Reagent 2: no characteristic odour

**pH:** Reagent 1 ( 30 °C ± 1 °C): 10.40  
Reagent 2 ( 25 °C ± 1 °C): 8.99  
R1+R2 ( 30 °C ± 1 °C): 10.40

**Absorbance:** Reagent 1: 0.000  
405 nm / 25 °C R1 + R2: 0.526

**Control sera:** TruLab N TruLab P  
Lot: 24382 Lot: 24383

Fresh (at 37°C) 82.4 U/L 208 U/L

Target values 77.5 58.1 - 96.9 199 149 - 249

---

**Additional remarks:** Mf-Date: 07.11.2018

This lot meets our specifications.  
This certificate of analysis is a computer printout and has therefore not been signed  
by hand.

Sign.: Dr. G. Gorka