

Quality Control

Certificate of Analysis

QC release date: 2020-05-06

Product: LDL-C Select FS
Product Code: 14121

Cat. No.: 1 4121 99 10 021

Delivery of kits: No of kits: 28
Kit Lot/Los: 60138594
Kit Expiry: 2021-08

Delivery of bulk: Volume:
Processing Lot:

Results for:

Reagent 1 Lot No.: 28433
Expiry: 2021-08-31

Reagent 2 Lot No.: 28433
Expiry: 2021-08-31

Examination:

Visual check Reagent 1: clear, yellowish to slightly pink
Reagent 2: clear, yellowish

Smell Reagent 1: no characteristic odour
Reagent 2: no characteristic odour

pH: Reagent 1 (25 °C ± 1 °C): 6.80
Reagent 2 (25 °C ± 1 °C): 7.00

Absorbance: Reagent 1: 0.002
600 - nm / 37 °C R1 + R2: 0.008

Control sera: TruLab L 1 TruLab L 2
Lot: 27037 Lot: 27038

Test results 94.5 mg/dL 151 mg/dL

Target values 93.6 74.9 - 112 151 121 - 181

Additional remarks: Mf-Date: 05.05.2020

This lot meets our specifications.

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka