

CERTIFICADO DE ANÁLISE CERTIFICATE OF ANALYSIS

| | |
|-------------------------------------|--|
| MS ANVISA | 80115310042 |
| CÓDIGO / CODE | 107095 |
| COD. INTERNO / INTERNAL CODE | 1100075K |
| DESCRIÇÃO / DESCRIPTION | LACTATO R1 3X20ML + R2 1X15ML |
| | KIT |
| | LOTE / LOT: 10709520123 |
| | DATA DE FABR./MANUF. DATE: 09/2020 |
| | DATA DE VALIDADE / EXPIRY DATE: 09/2021 |
| | REAGENTE 1 / REAGENT 1 |
| | COD. INTERNO / INTERNAL CODE: 1100020R1K |
| | LOTE / LOT: 10721320103 |
| | DATA DE VALIDADE / EXPIRY DATE: 09/2021 |
| | REAGENTE 2 / REAGENT 2 |
| | COD. INTERNO / INTERNAL CODE: 1100015R2K |
| | LOTE / LOT: 10721420103 |
| | DATA DE VALIDADE / EXPIRY DATE: 09/2021 |

| DESCRIÇÃO / DESCRIPTION | ESPECIFICAÇÃO / SPECIFICATION | RESULTADO / RESULTS |
|---------------------------------------|--|---------------------|
| APARENCIA/APPEARANCE R1 | LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH | Conforme/According |
| APARENCIA/APPEARANCE R2 | LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH | Conforme/According |
| ODOR/ODOUR R1 | ODOR CARACTERISTICO DETERGENTES - CHARACTERISTIC ODOUR DETERGENT | Conforme/According |
| ODOR/ODOUR R2 | SEM ODOR CARACTERISTICO - NO CHARACTERISTIC ODOUR | Conforme/According |
| pH (25 oC) R1 | 8.90 - 9.10 | Conforme/According |
| PH (25 oC) R2 | 2,50 - 2,70 | Conforme/According |
| pH (25 oC) R1 + R2 | 8.75 - 9.15 | Conforme/According |
| ABSORBANCIA/ABSORBANCE R1 - 340 nm | 0.000 - 0.100 | Conforme/According |
| ABSORBANCIA/ABSORBANCE R2 - 340 nm | 0,000 - 0,100 | Conforme/According |
| ABSORBANCIA/ABSORBANCE R1+R2 - 340 nm | 0.000 - 0.300 | Conforme/According |
| TRULAB N - LOTE: 27547 (mg/dL) | 9.01 - 13.80 | 12.63 |
| TRULAB P - LOTE: 27548 (mg/dL) | 24.60 - 37.60 | 31.5 |

OBS.: Este documento foi produzido eletronicamente e é válido sem assinatura / This document has been produced eletronicly and is valid without a signature.

Monique Brasil Vianna de Sá
Controle de Qualidade / Quality Control