

CERTIFICADO DE ANÁLISE CERTIFICATE OF ANALYSIS

MS ANVISA	80115310042	
CÓDIGO / CODE	107095	
COD. INTERNO / INTERNAL CODE	1100075K	
DESCRIÇÃO / DESCRIPTION	LACTATO R1 3X20ML + R2 1X15ML	
	KIT	
	LOTE / LOT:	10709521132
	DATA DE FABR./MANUF. DATE	06/2021
	DATA DE VALIDADE / EXPIRY DATE:	07/2022
	REAGENTE 1 / REAGENT 1	
	COD. INTERNO / INTERNAL CODE:	1100020R1K
	LOTE / LOT:	10721321109
	DATA DE VALIDADE / EXPIRY DATE:	07/2022
	REAGENTE 2 / REAGENT 2	
	COD. INTERNO / INTERNAL CODE:	1100015R2K
	LOTE / LOT:	10721421109
	DATA DE VALIDADE / EXPIRY DATE:	07/2022

DESCRIÇÃO / DESCRIPTION	ESPECIFICAÇÃO / SPECIFICATION	RESULTADO / RESULTS
APARENCIA/APPEARANCE R1	LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH	CONFORME/ACCORDING
APARENCIA/APPEARANCE R2	LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH	CONFORME/ACCORDING
ODOR/ODOUR R1	ODOR CARACTERISTICO DETERGENTES - CHARACTERISTIC ODOUR DETERGENT	CONFORME/ACCORDING
ODOR/ODOUR R2	SEM ODOR CARACTERISTICO - NO CHARACTERISTIC ODOUR	CONFORME/ACCORDING
pH (25 oC) R1	8.90 - 9.10	CONFORME/ACCORDING
PH (25 oC) R2	2,50 - 2,70	CONFORME/ACCORDING
pH (25 oC) R1 + R2	8.75 - 9.15	CONFORME/ACCORDING
ABSORBANCIA/ABSORBANCE R1 - 340 nm	0.000 - 0.100	CONFORME/ACCORDING
ABSORBANCIA/ABSORBANCE R2 - 340 nm	0,000 - 0,100	CONFORME/ACCORDING
ABSORBANCIA/ABSORBANCE R1+R2 - 340 nm	0.000 - 0.300	CONFORME/ACCORDING
TRULAB N - LOTE: 27547 (mg/dL)	9.01 - 13.80	11.84
TRULAB P - LOTE: 27548 (mg/dL)	24.60 - 37.60	31.22

OBS.: Este documento foi produzido eletronicamente e é válido sem assinatura / This document has been produced eletronically and is valid without a signature.

Monique Brasil Vianna de Sá
Controle de Qualidade / Quality Control