

Quality Control

## Certificate of Analysis

QC release date: 2018-01-23

**Product:** Total Protein UC FS  
**Product Code:** 10210

Cat. No.: 1 0210 99 10 021

---

**Delivery of kits:** No of kits: 10  
Kit Lot/Los: 60123320  
Kit Expiry: 2019-12

---

**Delivery of bulk:** Volume:  
Processing Lot:

---

### Results for:

Reagent Lot No.: 24674  
Expiry: 2019-12

### Examination:

Visual check Reagent : clear, orange-red  
Smell Reagent : no characteristic odour

**pH:** Reagent ( 25 °C ± 1 °C): 2.54

**Absorbance:** Reagent : 0.081  
600 nm/ 15-25°C

<b>Control sera:</b>	TruLab Urin 1	TruLab Urin 2
	Lot: 22728	Lot: 22729
Fresh (at 37°C)	0.089 g/L	0.673 g/L
Target values	0.086 0.066 - 0.107	0.647 0.492 - 0.803

---

### Additional remarks:

This lot meets our specifications.  
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka