

Quality Control

Certificate of Analysis

QC release date: 2019-11-04

Product: Total Protein UC FS
Product Code: 10210

Cat. No.: 1 0210 99 10 021

Delivery of kits: No of kits: 24
Kit Lot/Los: 60135366
Kit Expiry: 2021-08

Delivery of bulk: Volume:
Processing Lot:

Results for:

Reagent Lot No.: 27976
Expiry: 2021-10-20

Examination:

Visual check: Reagent : clear, orange-red
Smell Reagent : no characteristic odour
pH: Reagent (25 °C ± 1 °C): 2.48

Absorbance: Reagent : 0.081
600 nm/ 15-25°C

Control sera:	TruLab Urine 1	TruLab Urine 2
	Lot: 25444	Lot: 25445
Fresh (at 37°C)	0.079 g/L	0.632 g/L
Target values	0.082 0.063 - 0.102	0.644 0.490 - 0.799

Additional remarks: Mf-Date: 21.10.2019

This lot meets our specifications.
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka