

Quality Control

## Certificate of Analysis

QC release date: 2020-09-28

**Product:** Total Protein UC FS  
**Product Code:** 10210

Cat. No.: 1 0210 99 10 021

---

**Delivery of kits:** No of kits: 10  
Kit Lot/Los: 60141192  
Kit Expiry: 2022-03

---

**Delivery of bulk:** Volume:  
Processing Lot:

---

### Results for:

Reagent Lot No.: 29270  
Expiry: 2022-09-04

### Examination:

Appearance: Reagent : clear, orange-red  
Odour: Reagent : no characteristic odour  
**pH:** Reagent ( 25 °C ± 1 °C): 2.59  
**Absorbance:** Reagent : 0.119  
600 nm/ 15-25°C

<b>Control sera:</b>	TruLab Urine 1	TruLab Urine 2
	Lot: 27321	Lot: 27322
Fresh (at 37°C)	0.076 g/L	0.603 g/L
Target values	0.070 0.053 - 0.087	0.644 0.489 - 0.799

---

**Additional remarks:** Mf-Date: 04.09.2020

This lot meets our specifications.  
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka