

Quality Control

Certificate of Analysis

QC release date: 2020-12-11

Product: Total Protein UC FS
Product Code: 10210

Cat. No.: 1 0210 99 10 021

Delivery of kits: No of kits: 30
Kit Lot/Los: 60143529
Kit Expiry: 2022-03-27

Delivery of bulk: Volume:
Processing Lot:

Results for:

Reagent Lot No.: 29883
Expiry: 2022-11-20

Examination:

Appearance: Reagent : clear, orange-red
Odour: Reagent : no characteristic odour
pH: Reagent (25 °C ± 1 °C): 2.54
Absorbance: Reagent : 0.097
600 nm/ 15-25°C

Control sera:	TruLab Urine 1	TruLab Urine 2
	Lot: 27321	Lot: 27322
Fresh (at 37°C)	0.072 g/L	0.651 g/L
Target values	0.070 0.053 - 0.087	0.644 0.489 - 0.799

Additional remarks: Mf-Date: 20.11.2020

This lot meets our specifications.
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka