

Quality Control

Certificate of Analysis

QC release date: 2019-10-25

Product: Triglycerides FS
Product Code: 15710

Cat. No.: 1 5710 99 10 023

Delivery of kits: No of kits: 12
Kit Lot/Los: 60136708
Kit Expiry: 2021-10

Delivery of bulk: Volume:
Processing Lot:

Results for:

Reagent Lot No.: 27967
Expiry: 2021-10-13

Examination:

Visual check: Reagent : clear, yellow

Smell Reagent : like phenol

pH: Reagent (25 °C ± 1 °C): 7.20

Absorbance: Reagent : 0.017
546 nm/ 15-25°C

Time to reach endpoint ≤ 10 minutes at 37 °C

Control sera:	TruLab N Lot: 25894	TruLab P Lot: 25895
Fresh (at 37°C)	92.5 mg/dL	173 mg/dL
Target values	92.9 76.2 - 110	171 141 - 202

Additional remarks:

This lot meets our specifications.
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka