

Quality Control

Certificate of Analysis

QC release date: 2020-05-07

Product: Triglycerides FS
Product Code: 15710

Cat. No.: 1 5710 99 10 023

Delivery of kits: No of kits: 60
Kit Lot/Los: 60140733
Kit Expiry: 2022-04

Delivery of bulk: Volume:
Processing Lot:

Results for:

Reagent Lot No.: 28811
Expiry: 2022-04-24

Examination:

Appearance: Reagent : clear, yellow

Odour: Reagent : like phenol

pH: Reagent (25 °C ± 1 °C): 7.16

Absorbance: Reagent : 0.021
546 nm/ 15-25°C

Time to reach endpoint ≤ 10 minutes at 37 °C

Control sera:	TruLab N Lot: 25894	TruLab P Lot: 25895
Fresh (at 37°C)	91.4 mg/dL	172 mg/dL
Target values	92.9 76.2 - 110	171 141 - 202

Additional remarks: Mf-Date: 22.04.2020

This lot meets our specifications.
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka