

DiaSys Diagnostic Systems GmbH
65558 Holzheim, GERMANY

Quality Control

Certificate of Analysis

QC release date: 2020-03-31

Product: TruLab CRP hs Level 2
Product Code: 59740

Cat. No.: 5 9740 99 10 046

Delivery of kits: No of kits: 2
Kit Lot/Los: 60140579
Kit Expiry: 2022-03

Results for:

Control Lot No.: 28981
Expiry: 2022-03-06

Assay value	Range	Units
--------------------	--------------	--------------

1.96	1.43 - 2.49	mg/L
------	-------------	------

Description: This Control is manufactured from a human serum base and all the donors are individually tested and shown using FDA-approved methods to be free from HBsAG, HCV and HIV antibodies.

Additional remarks: Mf-Date: 06.03.2020

This lot meets our specifications.
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka