

Quality Control

Certificate of Analysis

QC release date: 2017-05-23

Product: Urea FS
Product Code: 13101

Cat. No.: 1 3101 99 10 920

Delivery of kits: No of kits: 3
Kit Lot/Los: 60118173
Kit Expiry: 2019-05

Delivery of bulk: Volume:
Processing Lot:

Results for:

Reagent 1 Lot No.: 23625
Expiry: 2019-05

Reagent 2 Lot No.: 23625
Expiry: 2019-05

Examination:

Visual check Reagent 1: clear, colourless
Reagent 2: clear, pale yellow

Smell Reagent 1: slight smell (Thiol)
Reagent 2: no characteristic odour

pH: Reagent 1 (25 °C ± 1 °C): 7.82
Reagent 2 (25 °C ± 1 °C): 9.60
R1+R2 (25 °C ± 1 °C): 7.81

Absorbance: Reagent 1: 0.236
340 nm / 15-25 °C R1 + R2: 1.932

Control sera: TruLab N TruLab P
Lot: 21591 Lot: 21592

Fresh (at 37°C) 37.6 mg/dL 154 mg/dL

Target values 38.9 30.3 - 47.5 154 120 - 188

Additional remarks:

This lot meets our specifications.
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka