

Quality Control

Certificate of Analysis

QC release date: 2019-04-08

Product: Lp(a) 21 FS
Product Code: 17139

Cat. No.: 1 7139 99 10 930

Delivery of kits: No of kits: 3
Kit Lot/Los: 60133866
Kit Expiry: 2021-03

Delivery of bulk: Volume:
Processing Lot:

Results for:

Reagent 1 Lot No.: 27044
Expiry: 2021-03

Reagent 2 Lot No.: 27044
Expiry: 2021-03

Examination:

Visual check Reagent 1: clear to slightly opalescent,
yellowish
Reagent 2: milky, white

Smell Reagent 1: no characteristic odour
Reagent 2: no characteristic odour

Control sera: TruLab Lp(a) 1 TruLab Lp(a) 2
Lot: 26192 Lot: 26193

Test results 42.7 nmol/L 106 nmol/L

Target values 39.7 31.8 - 47.6 105 84.0 - 126

Additional remarks: Mf-Date: 11.03.2019

This lot meets our specifications.

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka